



#### KEY NOTES

1. QUANTITY OF X5320, FILM PROCESSOR, MAY VARY ACCORDING TO WORK LOAD.
2. P6150, SINK, IS REQUIRED ONLY FOR THRU WALL FILM PROCESSOR SYSTEMS, X5320. THIS IS TO BE LOCATED OUTSIDE THE PROCESSING ROOM AND WITHIN PROXIMITY TO THE X5320.

#### General Notes:

1. For typical equipment mounting heights see "TYPICAL EQUIPMENT MOUNTING HEIGHTS" drawing.
2. MIL-STD-1691 data can only be accessed from the Equipment Plan.

## XRAY, MAMMO PROCESSING (XDMP1)

RADIOLOGY/NUCLEAR MEDICINE

ELEVATIONS

9.75 NSM (105 NSF)

DATE: 07/12/2001

0 1000 2000



1:50 SCALE (MILLIMETERS)